

Corrigendum

No. IITMANDI/S&P/PUR-203/2017-18/8349-50

Date: 06.12.2017

To

1. Institute Website IIT Mandi
2. Central Public Procurement Portal.

Sub: -Corrigendum to IITMANDI/S&P/PUR-203/2017-18/7752-53

Tender date: 15th November, 2017 for the supply of “GLOVE BOX”.

The following addition & changes are made in the specification of the Tender

GLOVE BOX.

Sr.No.	Mentioned in the Tender	May be read as
1.	Inside dimensions : 1200 x 600 x 600 mm ³ or Larger	1. Inside dimensions: 1200 x 600 x 600 mm ³ or Larger preferable
2.	Window Chemical resistant and inclined panel of 9-10 mm thickness with Polycarbonate window on front	1. Window: Chemical resistant and inclined panel of 7-10 mm thickness with Polycarbonate or Chemical resistant any appropriate toughened glass etc. window on front
3.	Lightning Internally mounted 2 LED Lamps	1. Lightning: Internally mounted 2 LED Lamps or externally fluorescent suitable light
4.	Ante Chambers A) Inside Dimensions (1) 500 mm (D) x 500 mm (H) (2) Gas lines with dual control valves must be given for purging purposes.	2. Ante Chambers: a) Large Ante Chamber: Inside Dimensions (1) 350 mm (Higher preferable) (D) x 550 mm (L) (Higher preferable) (2) Gas lines with dual control valves must be given for purging purposes (automatically purging with electro pneumatic/solenoid valves preferable) b) Small Ante Chamber: (1) Dimensions; 120 mm (D) x 300 mm (L) (2) Gas lines with dual control valves must be given for purging purposes
5.	Oxygen Sensors 0.001-1000 ppm with solvent trap	Oxygen sensor: 0.01-1000 ppm with solvent trap

6.		1. Purity Level: < 1ppm (Lower is preferable) O ₂ , H ₂ O in Nitrogen, Argon or Helium etc.
7.		2. Purifier capacity: 50L (higher preferable) for Oxygen and 1 Kg (higher preferable) for Moisture with auto regeneration.
8		3. Purifier Materials: a) Oxygen Catalyst b) Molecular Sieves c) Solvent Adsorbent – Activated Carbon
9.		High efficiency and low noise circulation blower with flow rate: 40 m ³ /hr (Higher preferable), flow control by oxygen level in the chamber

Those who have already uploaded their bids as per earlier terms and conditions can upload their revised bids in the light of the correct revised specifications as mentioned above. Rest of the specifications, terms and conditions will remain the same.

The Last date for submission of tender has been extended from **5th December, 2017 to 14th December, 2017 till 03:00 P.M. and date for Technical Bid opening has been extended from 6th December, 2017 to 15th December, 2017 at 4:00 PM.**

For any technical quires and clarification regarding the tender, the bidders are requested to contact in office through e-mail: satinder@iitmandi.ac.in.

Sd/-
**Assistant Registrar
(Stores & Purchase)**