
CORRIGENDUM

No. IIT Mandi/Admin/PUR-133/2014-15/7522-23

Date: 11.02.2015

To

1. Institute Website IIT Mandi
2. Central Public Procurement Portal.

Sub: - Corrigendum to NIT No. IITMANDI/Admin/PUR-133/2014-15/7221-22 dated 29.01.2015 regarding Supply & Installation of Raman Spectrometer.

The **Thermal Stability** has been mentioned in Annexure 1 and 2 regarding technical specification for **Heating Stage (above 300 K) and Close Cycle Refrigerator (CCR)** for Raman and Photo-Luminescent measurements and Raman Imaging at **may be read as 100 mK (or better) instead of 10 mK (or better) at 4K (for CCR).**

Those who have already sent their bids as per earlier technical specification can send their revised bids in light of the correct technical specification as mentioned above. However, they need not submit bid fee & EMD again.

All other terms and conditions will remain the same.

Sd/-
Assistant Registrar
(Stores & Purchase)