

CORRIGENDUM

No.: IITMANDI/S&P/PUR-232/2015-16/2022-23

Dated:10.06.2016

To

1. Institute Website IIT Mandi
2. Central Public Procurement Portal.

Sub: - Corrigendum to NIT No. IITMANDI/S&P/PUR-232/2015-16/1818 -19 dated 03rd June, 2016 regarding Differential Scanning Calorimeter (DSC).

The following changes are made in the specification of the **Differential Scanning Calorimeter (DSC)**.

Sr. No.	Mentioned in the Tender	May be read as
1.	Annexure A: Point No. 6 Feature Calorimetric Sensitivity 0.1 μ W or better	Calorimetric sensitivity 1.0 μ W

Those who have already sent their bids as per earlier technical specification can send their revised bids in the light of the correct technical specification as mentioned above.

All other terms and conditions will remain the same.

Sd/-
**(Stores & Purchase)
Assistant Registrar**