

Corrigendum

No. IITMANDI/S&P/PUR-200/2017-18/8351-52

Date: 06.12.2017

To

1. Institute Website IIT Mandi
2. Central Public Procurement Portal.

Sub: -Corrigendum to IITMANDI/S&P/PUR-200/2017-18/7743-44

Tender date: 15th November, 2017 for the supply of “Field Emission Scanning Electron Microscope (FESEM)”.

The following changes are made in the specification of the Tender **Field Emission Scanning Electron Microscope (FESEM)**.

Sr.No.	Mentioned in the Tender	May be read as
1.	Sr. No. 3 : Resolution:- 0.7nm or below at 15kV and 1.2 nm or below at 1kV	Sr. No. 3 : Resolution:- 0.7nm or better at 15kV and 1 nm or better at 1kV
2.	Sr. No. 10 : Stage 5 axis motorized eccentric stage with motorized stage movements equivalent to or better X=100 mm or more Y=100 mm or more Z=50mm or more Tilt=-3-70° Rotation=360° ; should be continuously varying	Sr. No. 10 : Stage 5 axis motorized eccentric/ compucentric stage with motorized stage movements Equivalent to or better X=100 mm or more Y=100 mm or more Z=50mm or more Tilt=-3-70° Rotation=360° ; should be continuously varying (Full detail of tilt and Rotation specifications should be provided)
3.	Sl. No. 11. Detectors:- a) Everhart-Thronley (ET) Secondary Electron Detector. b) In-Lens /In column SE detector detector/upper secondary electron detector or equivalent detector for high resolution imaging in High Vacuum with automatic brightness & contrast adjustment. c) Parallel in-column detection of secondary electrons and backscattered electrons.	Sr. No. 11. Detectors:- a) Secondary Electron Detector . b) In-Lens/In column SE detector detector/upper secondary electron detector or equivalent detector for high resolution imaging in High Vacuum with automatic brightness & contrast adjustment. c). Parallel in-column detection of secondary electrons and backscattered electrons
4.	Sr. No. 15. Compressor :- For all pneumatic operations	Sr. No. 15. Compressor: For all pneumatic operations (If required)

Those who have already uploaded their bids as per earlier terms and conditions can upload their revised bids in the light of the correct revised specifications as mentioned above. Rest of the specifications, terms and conditions will remain the same.

The Last date for submission of tender has been extended from **5th December, 2017 to 14th December, 2017 till 03:00 P.M. and date for Technical Bid opening has been extended from 6th December, 2017 to 15th December, 2017 at 3:00 PM.**

For any technical quires and clarification regarding the tender, the bidders are requested to contact in office through e-mail: satinder@iitmandi.ac.in.

Sd/-
**Assistant Registrar
(Stores & Purchase)**